

Technical Data Sheet

Pulse Vacuum Steam Sterilizer

Model: ICLAVE-350SD-B



The picture is for reference only, which shall not be taken as standard for machine acceptance. For details, it is subject to technical description.

1 Technical data

The system has **32 built-in preset programs**, more than 10 kinds of program stage can be flexibly configured according to the needs to meet the requirements of different sterilization process.

Name	Ster. Temp.	Ster. Time	Dry time	Applicable items type		
Warm up	134°C	0min	3min	Empty loading, for preheat the device		
B&D test	134°C	3.5min	4min	B&d test package or a device		
Fabric	134°C	5min	10min	Fabric package, weight ≤7.5kg/pack		
Instrument	134°C	5min	15min	Conventional instrument box or basket loading, weight≤ 7.5kg/pack		
Thermolabile	121℃	20min	15min	Items can't bear 134°C, weight ≤ 7.5kg/pack		
Flash	134°C	4min	1min	Unwrapped instruments		
Orthopedics	134℃	6min	15min+10min	Orthopedics instruments, weight ≤14kg/pack		
Leak test				Empty loading, leak rate ≤ 0.13kpa/min		
Prion	134°C	30min	15min	Special items such as prions		
Optical	134°C	7min	15min+5min	Inner diameter ≥ 2mm, length ≤ 1500 times in diameter from the opening side to end		
Heavy load	134°C	6min	15min+10min	Heavy loading items		
Small load	134°C	5min	8min	loading capacity < one standard sterilization ur weight ≤7.5kg/pack		
Open liquid	121℃	30min		Unsealed bottled liquid, volume ≤ 500ml / bottle		
Gravity	121℃	20min		Gravity steam discharge, non-vacuum		

1. Sterilizing program

		Positive	134°C	7	7min	10min	ı	Items needs positive pressure replacement and vacuum drying		
2.	Designed pressure					-0.1~0.3 MPa				
	Rated working pressure					0.25 MPa				
	Vacuum low limit				-0.09 MPa					
	Vacuum pulses counts				0~99 Times					
	Designed temperature					150°C				
						134°C				
	<u> </u>									
	<u> </u>					139°C				
_					Rectangular mm / 352Liters					
_	` '				mm / 332Liters					
	` '				1080K	σ				
	-				ಕ ation on th	o ara	ound			
_	Installation form Door opening method			IIIStalia	מנוטוו טוו נו	ie git	Suna			
					Double	De blades a				
	Quantity of doors				Double doors					
	Door opening direction			 	Up and down					
		or sealing method ntroller and screen			By compressed air with a door gasket sealing					
				Front side 8" inch color touch screen, 5 level authorization user						
		•		_	0.3~0.5MPa		Must be pure water, 0.04m³/cycle			
		'		5~0.3 MPa Soft water, 0.5m³/cycle						
21.	Cleaned com	<u>'</u>		Oil free & water free						
22.	Power supply	ZOWAR STINNIV			iving power: AC 380 V ± 10% 3 phases 50 Hz ntrol power: AC 220V ± 10% single phase 50Hz					
23.	Pressure disc				alogue pressure gauge for chamber and jacket on the front panel					
		· · ·			gital Temperature Display on the front panel (touch screen)					
	· · · · · · · · · · · · · · · · · · ·				kW, working pressure: 0.3MPa					
		ings on huilt-in steam generator			fety valve, Analogue pressure gauge, Water level gauge, water level ectric cut out, Automatic air ventilate, Control Switch					
27.	Water pump	ater numn:			ot less than 3 bar should be fitted with a protection against overload and ase failure.					
28.	Data record Bui			ilt-in micro printer						
29.			Internal loading cart x 1 and External transfer trolley x 2							
30.	Safety system Ove			er pressure protect, water level protect, door cannot open in case of essure, door obstacle system, overload protect, and alarming system						
31.	certification			ME	DD 93/42 EEC ISO ASMA EMC					
2	Component material									
	Component				Material					
1.	Chamber 9			SUS304						
2.				SUS304						
3.	Chamber insulation material			Rock wool						
4.	Chamber insulation cover			Embossed aluminum sheet						
				SUS304						
	Door plate 5			SUS304						
				Silicon rubber						
				Carbon steel						
					SUS304					
					SUS304					

11.	Internal loading ca	ırt		SUS304				
12.				SUS304				
3	· · · · · · · · · · · · · · · · · · ·							
No.		Model	Brand	Remark				
1.	Main Chamber body	XG1.HW.01	MRC	Class I pressure vessel. Welding by robot Inner chamber is adopted 304 stainless steel; The jacket is 304 steel.				
2.	Door	XG1.HW.03	MRC	The door inner face is adopted 304 stainless steel; It is electric sliding and compressed air sealing, equipped with safety interlocking and manually controlled open equipment.				
3.	Door control switch	MLCA12-TH	OMRON Japan	Operating Reliably, heat resistant, long service life				
4.	PLC	XPC39160	MRC	Strong function, advanced performance, high reliability, Multi communication mode.				
5.	Touch screen	NSC08-60	MRC	8 inch color touch screen, display working process parameters, easy control and operation. Select Program to Run P-Chamber 1.0 kPa P-Jacket 210 kPa T-Chamber 25.0 To T-Jacket 133.5 To T-Jack				
6.	Sterilization software	Wincc flexible	MRC	Multi programs, program modularization management.				
7.	thermal printer	WH4008A	MRC	Core made in Japan; multi record channels, sterilizing parameter record, long service life.				
8.	Pressure transmitter	ECO-1-ABS	WIKA, German					
9.	Pressure gauge	-0.1~0.4MPa	WIKA , German	High precision				
10.	Temperature sensor	Pt100	WIKA , German	High precision, mini-measurement error.				
11.	Pneumatic valve	554 series	GEMU, German	Powerful switch valve, no action error, remote compressed air				
12.	Vacuum pump	2BV series	Nash Elmo, America	Running stable, no water leakage, high reliability, low noisy.				
13.	Air filter	CHL0.2	MRC	Ultra-fine sterile filtration, bacterial eliminating rate ≥99.97%				
14.	Safety valve	0.3MPa	MRC					
15.								
4	Programs (some	of)						





